

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 06/20/2024

RMS
24-27551

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:57
INTERNAL STANDARD	VERIFIED	08:58
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:58
BLANK TEST	0.00	08:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:59
BLANK TEST	0.00	09:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:00
BLANK TEST	0.00	09:01

