

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 07/02/2024

*2MS
24-29537*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:47
INTERNAL STANDARD	VERIFIED	13:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:47
BLANK TEST	0.00	13:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:48
BLANK TEST	0.00	13:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:50
BLANK TEST	0.00	13:50

