

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 07/30/2024

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 24140

*2MS  
 24-33707*

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:50
INTERNAL STANDARD	VERIFIED	10:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:51
BLANK TEST	0.00	10:52
REFERENCE STANDARD TEMP	33.98c	
REFERENCE STANDARD	0.09	10:52
BLANK TEST	0.00	10:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:53
BLANK TEST	0.00	10:54

