

SOLUTION CHANGE

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 02/01/2024

*RM*  
*21-5218*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36882  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:58
INTERNAL STANDARD	VERIFIED	11:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:59
BLANK TEST	0.00	12:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:00
BLANK TEST	0.00	12:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:01
BLANK TEST	0.00	12:02

