

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 02/06/2024

LMS
24-5563

OPERATOR NAME:
MICHAEL J. CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:30
INTERNAL STANDARD	VERIFIED	14:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	14:30
BLANK TEST	0.00	14:31
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:32
BLANK TEST	0.00	14:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:33
BLANK TEST	0.00	14:33

