

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RM>
24-11304

Date: 03/14/2024

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:52
INTERNAL STANDARD	VERIFIED	10:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:52
BLANK TEST	0.00	10:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:53
BLANK TEST	0.00	10:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:54
BLANK TEST	0.00	10:55

