

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

Date: 05/07/2024

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 24040

RMS
24-20013

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:11
INTERNAL STANDARD	VERIFIED	09:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:11
BLANK TEST	0.00	09:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:12
BLANK TEST	0.00	09:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:14
BLANK TEST	0.00	09:14

