

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 07/09/2024

RMS
24-30552

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:30
INTERNAL STANDARD	VERIFIED	09:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:31
BLANK TEST	0.00	09:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:32
BLANK TEST	0.00	09:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:33
BLANK TEST	0.00	09:34

