

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 07/16/2024

*RMS
24-31581*

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36602
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:36
INTERNAL STANDARD	VERIFIED	12:36
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:37
BLANK TEST	0.00	12:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:38
BLANK TEST	0.00	12:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:39
BLANK TEST	0.00	12:40

