

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

12ms
24-32614

Date: 07/23/2024

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:53
INTERNAL STANDARD	VERIFIED	09:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:53
BLANK TEST	0.00	09:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:55
BLANK TEST	0.00	09:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:56
BLANK TEST	0.00	09:56

