

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 02/04/2025

RMJ
25-5387

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:26
INTERNAL STANDARD	VERIFIED	09:26
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:27
BLANK TEST	0.00	09:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:28
BLANK TEST	0.00	09:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:29
BLANK TEST	0.00	09:30

