

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/15/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:43
INTERNAL STANDARD	VERIFIED	09:43
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:43
BLANK TEST	0.00	09:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:44
BLANK TEST	0.00	09:45
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:45
BLANK TEST	0.00	09:46

RMS
25-16325

