

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/29/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:03
INTERNAL STANDARD	VERIFIED	09:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:03
BLANK TEST	0.00	09:04
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:05
BLANK TEST	0.00	09:06
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:06
BLANK TEST	0.00	09:07

