

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 05/06/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25040

RLS
25-19807

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:11
INTERNAL STANDARD	VERIFIED	09:11
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	09:12
BLANK TEST	0.00	09:12
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:13
BLANK TEST	0.00	09:14
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	09:14
BLANK TEST	0.00	09:15

