

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 08/26/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:59
INTERNAL STANDARD	VERIFIED	09:59
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:59
BLANK TEST	0.00	10:00
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:00
BLANK TEST	0.00	10:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:02
BLANK TEST	0.00	10:02

