

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 11/03/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:21
INTERNAL STANDARD	VERIFIED	10:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:22
BLANK TEST	0.00	10:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:23
BLANK TEST	0.00	10:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:24
BLANK TEST	0.00	10:25

