

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 11/25/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:53
INTERNAL STANDARD	VERIFIED	10:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:53
BLANK TEST	0.00	10:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:55
BLANK TEST	0.00	10:55
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	10:56
BLANK TEST	0.00	10:57

RMS
25-52711

