

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 05/01/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:05
INTERNAL STANDARD	VERIFIED	09:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:05
BLANK TEST	0.00	09:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:06
BLANK TEST	0.00	09:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:08
BLANK TEST	0.00	09:08

