

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 05/27/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:16
INTERNAL STANDARD	VERIFIED	09:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:17
BLANK TEST	0.00	09:17
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:18
BLANK TEST	0.00	09:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:19
BLANK TEST	0.00	09:20

