

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 10/21/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:03
INTERNAL STANDARD	VERIFIED	09:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:04
BLANK TEST	0.00	09:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:05
BLANK TEST	0.00	09:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:06
BLANK TEST	0.00	09:07

Rms 25-47703

