

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 10/29/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:46
INTERNAL STANDARD	VERIFIED	09:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:46
BLANK TEST	0.00	09:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:48
BLANK TEST	0.00	09:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:49
BLANK TEST	0.00	09:50

25-4888

