

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 11/25/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:02
INTERNAL STANDARD	VERIFIED	11:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	11:02
BLANK TEST	0.00	11:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:03
BLANK TEST	0.00	11:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:04
BLANK TEST	0.00	11:05

