

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 12/01/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:44
INTERNAL STANDARD	VERIFIED	09:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	09:44
BLANK TEST	0.00	09:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:45
BLANK TEST	0.00	09:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:46
BLANK TEST	0.00	09:47

RMS
25-53631

