

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 12/02/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:33
INTERNAL STANDARD	VERIFIED	09:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:33
BLANK TEST	0.00	09:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:34
BLANK TEST	0.00	09:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:35
BLANK TEST	0.00	09:36

RMS
25-54/72

