

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 12/16/2025

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36882
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:33
INTERNAL STANDARD	VERIFIED	08:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:34
BLANK TEST	0.00	08:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:35
BLANK TEST	0.00	08:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:36
BLANK TEST	0.00	08:37

RM
25-55869

