

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 10/05/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 38303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26270

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:26
INTERNAL STANDARD	VERIFIED	15:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:27
BLANK TEST	0.00	15:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:28
BLANK TEST	0.00	15:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	15:29
BLANK TEST	0.00	15:30

26-40842

