

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 10/13/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 38303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26270

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:43
INTERNAL STANDARD	VERIFIED	08:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:43
BLANK TEST	0.00	08:44
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:44
BLANK TEST	0.00	08:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:46
BLANK TEST	0.00	08:46

26-41554

