

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308




Date: 04/14/2026

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:13
INTERNAL STANDARD	VERIFIED	08:13
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:13
BLANK TEST	0.00	08:14
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:14
BLANK TEST	0.00	08:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:15
BLANK TEST	0.00	08:16

26-14237



SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 04/14/2026

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26050

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:08
INTERNAL STANDARD	VERIFIED	08:09
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:09
BLANK TEST	0.00	08:10
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:10
BLANK TEST	0.00	08:11
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	08:11
BLANK TEST	0.00	08:12

26-14237

