

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308


Date: 08/03/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 388303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:56
INTERNAL STANDARD	VERIFIED	09:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:56
BLANK TEST	0.00	09:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:57
BLANK TEST	0.00	09:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:58
BLANK TEST	0.00	09:59

26-31147



SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 08/03/2026

OPERATOR NAME:
JOSHUA R HARRINGTON
Permit #: 38303
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:49
INTERNAL STANDARD	VERIFIED	09:49
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	09:50
BLANK TEST	0.00	09:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:51
BLANK TEST	0.00	09:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:52
BLANK TEST	0.00	09:53

26-31147

