

SOLUTION CHANGE

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 06/01/2021

OPERATOR NAME: *RM'*
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21100

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:36
INTERNAL STANDARD	VERIFIED	13:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:36
BLANK TEST	0.00	13:37
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:37
BLANK TEST	0.00	13:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:39
BLANK TEST	0.00	13:39

