

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 02/09/2021

OPERATOR NAME:
MICJHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 20500

*Rm 5
21-4910*

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:28
INTERNAL STANDARD	VERIFIED	12:28
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:28
BLANK TEST	0.00	12:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:29
BLANK TEST	0.00	12:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:30
BLANK TEST	0.00	12:31

