

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

Date: 03/09/2021

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 20500

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:21
INTERNAL STANDARD	VERIFIED	11:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:21
BLANK TEST	0.00	11:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:23
BLANK TEST	0.00	11:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:24
BLANK TEST	0.00	11:24

*RM S*  
*21 8454*

