

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RMS
21-1/01

Date: 01/04/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 20430

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:23
INTERNAL STANDARD	VERIFIED	13:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:24
BLANK TEST	0.00	13:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:25
BLANK TEST	0.00	13:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	13:26
BLANK TEST	0.00	13:27

