

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 11/23/2021

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 366682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21280

ZMS
21-1/6/89

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:51
INTERNAL STANDARD	VERIFIED	12:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:52
BLANK TEST	0.00	12:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:54
BLANK TEST	0.00	12:55

