

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 02/09/2022

RMJ
22-49811

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:22
INTERNAL STANDARD	VERIFIED	14:22
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:22
BLANK TEST	0.00	14:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:23
BLANK TEST	0.00	14:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:25
BLANK TEST	0.00	14:25

