

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

*RMS*  
*22-6600*

Date: 02/22/2022

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:22
INTERNAL STANDARD	VERIFIED	15:23
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:23
BLANK TEST	0.00	15:24
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:24
BLANK TEST	0.00	15:25
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	15:25
BLANK TEST	0.00	15:26

