

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

ZMS
22-9400

Date: 03/16/2022

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:19
INTERNAL STANDARD	VERIFIED	13:19
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:19
BLANK TEST	0.00	13:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:21
BLANK TEST	0.00	13:21
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:22
BLANK TEST	0.00	13:22

