

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

Date: 03/29/2022

*RM's*  
*22-11167*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 21440

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:50
INTERNAL STANDARD	VERIFIED	12:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:51
BLANK TEST	0.00	12:51
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:52
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:54

