

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 05/24/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:17
INTERNAL STANDARD	VERIFIED	12:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:18
BLANK TEST	0.00	12:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:19
BLANK TEST	0.00	12:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:20
BLANK TEST	0.00	12:21

RM)
22-19468

