

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 08/03/2022

EMs
22-30705

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22190

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:36
INTERNAL STANDARD	VERIFIED	14:36
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:37
BLANK TEST	0.00	14:38
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:38
BLANK TEST	0.00	14:39
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:39
BLANK TEST	0.00	14:40

