

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 100282

Date: 12/06/2022

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 22390

RMS
22-19213

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:17
INTERNAL STANDARD	VERIFIED	14:17
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	14:18
BLANK TEST	0.00	14:18
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:19
BLANK TEST	0.00	14:20
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	14:20
BLANK TEST	0.00	14:21

