

SOLUTION CHANGE

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RM>
22-25491

Date: 07/01/2022

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:29
INTERNAL STANDARD	VERIFIED	13:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:30
BLANK TEST	0.00	13:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:31
BLANK TEST	0.00	13:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:32
BLANK TEST	0.00	13:33

