

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 107308

*2MS*  
*22-29518*

Date: 07/26/2022

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 22100

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:18
INTERNAL STANDARD	VERIFIED	13:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:18
BLANK TEST	0.00	13:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:19
BLANK TEST	0.00	13:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	13:20
BLANK TEST	0.00	13:21

