

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 100282

*RMS*  
*23-2591*

Date: 01/17/2023

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:01
INTERNAL STANDARD	VERIFIED	13:01
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:01
BLANK TEST	0.00	13:02
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:02
BLANK TEST	0.00	13:03
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:04
BLANK TEST	0.00	13:04

