

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

RMS
23-11247

Date: 03/14/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:47
INTERNAL STANDARD	VERIFIED	12:47
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:48
BLANK TEST	0.00	12:48
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:49
BLANK TEST	0.00	12:50
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	12:50
BLANK TEST	0.00	12:51

