

SUPERVISOR TEST

UTICA CITY  
 POLICE DEPT  
 DATAMASTER dmt: 100282

Date: 10/10/2023

*RMS*  
*23-46144*

OPERATOR NAME:  
 MICHAEL J CLIVE  
 Permit #: 36682  
 Operator Agency/Dept: UTICA CITY PD  
 Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:52
INTERNAL STANDARD	VERIFIED	13:52
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:52
BLANK TEST	0.00	13:53
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.09	13:53
BLANK TEST	0.00	13:54
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	13:55
BLANK TEST	0.00	13:55

