

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 100282

Date: 12/12/2023

LMJ
23-55397

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:29
INTERNAL STANDARD	VERIFIED	12:29
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:29
BLANK TEST	0.00	12:30
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:30
BLANK TEST	0.00	12:31
REFERENCE STANDARD TEMP	33.99c	
REFERENCE STANDARD	0.10	12:32
BLANK TEST	0.00	12:32

