

SUPERVISOR TEST

UTICA CITY  
POLICE DEPT  
DATAMASTER dmt: 107308

Date: 03/14/2023

*RMS  
23-11217*

OPERATOR NAME:  
MICHAEL J CLIVE  
Permit #: 36682  
Operator Agency/Dept: UTICA CITY PD  
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:56
INTERNAL STANDARD	VERIFIED	12:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:57
BLANK TEST	0.00	12:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:58
BLANK TEST	0.00	12:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:59
BLANK TEST	0.00	13:00

