

SUPERVISOR TEST

UTICA CITY
POLICE DEPT
DATAMASTER dmt: 107308

Date: 04/11/2023

OPERATOR NAME:
MICHAEL J CLIVE
Permit #: 36682
Operator Agency/Dept: UTICA CITY PD
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:56
INTERNAL STANDARD	VERIFIED	10:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:57
BLANK TEST	0.00	10:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:58
BLANK TEST	0.00	10:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:59
BLANK TEST	0.00	11:00

2ms
23-15715
Work Station

