

SUPERVISOR TEST

UTICA CITY
 POLICE DEPT
 DATAMASTER dmt: 107308

RMS
23-38225

Date: 08/15/2023

OPERATOR NAME:
 MICHAEL J CLIVE
 Permit #: 36682
 Operator Agency/Dept: UTICA CITY PD
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:22
INTERNAL STANDARD	VERIFIED	12:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:22
BLANK TEST	0.00	12:23
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:23
BLANK TEST	0.00	12:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:24
BLANK TEST	0.00	12:25

